

Configuration Guide | addition of Analog Manometer

An analog manometer may be added to the breathing circuit on the expiratory limb when additional verification of delivered PEEP is desired.

This is most useful when delivering pressure with devices that do not electronically control for pressure, such as wall-gas titration or venturi techniques.

